

Chloride**56I700190****20 - 12000 mg/L Cl⁻**

Material

Reagents	Packaging Unit	Part Number
Chloride LR Titrant CC2	65 mL	56L014265
Chloride HR Titrant BC2	65 mL	56L014165
Chloride Indicator BC1/CC1	65 mL	56L714065

The following accessories are required.

Accessories	Packaging Unit	Part Number
Syringe, plastic, 20 mL	1 Pieces	56A006501
Titration jar with cap, plastic, 60 mL	1 Pieces	56A006701
Syringe, plastic, 5 mL	1 Pieces	56A008501

Application List

- Cooling Water
- Boiler Water


Remarks

1. Alkaline samples such as boiler water will require neutralisation prior to testing.
2. Colours may vary depending on sample and test conditions.
3. Dilute samples of less than 10 mL to approximately 10-20 mL with distilled or deionised (chloride free) water.


Sampling

Select the sample volume from the table according to the expected measuring range and read off the factor to calculate the result.


Expected Range	Titrant used	Sample Size	Factor
20-75 mg/L	Chloride LR Titrant CC2	40 mL	2.5
50-150 mg/L	Chloride LR Titrant CC2	20 mL	5
100-400 mg/L	Chloride LR Titrant CC2	10 mL	10
100-400 mg/L	Chloride HR Titrant BC2	40 mL	10
200-600 mg/L	Chloride HR Titrant BC2	20 mL	20
400-1000 mg/L	Chloride HR Titrant BC2	10 mL	40
800-3000 mg/L	Chloride HR Titrant BC2	5 mL ³	80
2000-6000 mg/L	Chloride HR Titrant BC2	2 mL ³	200
4000-12000 mg/L	Chloride HR Titrant BC2	1 mL ³	400



Attention! Select the appropriate sample volume according to the instructions in the chapter Sampling.




Add **10** drops of **Chloride Indicator BC1/CC 1 (Potassium Chromate)** to give a **yellow** colour.



Attention! Record the number of drops that will be added.

Note: Make sure to swirl the jar after adding each drop!



Add **Chloride LR Titrant CC2** or **Chloride HR**

Titrant BC2 drop by drop to the sample until colouration turns from **yellow** to **orange/brown**.

Calculate test result: Chloride (as Cl) mg/L = Number of drops x factor (see table)