

# Lovibond® Colour Measurement Tintometer® Group



## FAC

Part Number: 131043



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### FAC

A single glass conformance filter for checking PFXi instruments in-FAC scale. This filter was previously provided as part of the accessories kit for the PFX195/3.

### Technical Data

Glass Standard Values	Approximate values
-----------------------	--------------------

Filters	1
---------	---

### Delivery Scope

- FAC 9 filter
- Certificate of conformity