

Lovibond® Colour Measurement Tintometer® Group



Lot de filtres Klett (filtre des bleus KS-42)

À utiliser uniquement avec le PFXi 880, 950, 995

Code: 139709



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Lot de filtres Klett (filtre des bleus KS-42)

A set of five glass conformance filters for checking Klett scale on PFXi instruments.

Fiche technique

Glass Standard Values	Valeurs approximatives : 21, 66, 162, 318, 616
Filters	5

Contenu de livraison

- Klett 21 filter
- Klett 66 filter
- Klett 162 filter
- Klett 318 filter
- Klett 616 filter
- Certificate of conformity