

# Lovibond® Colour Measurement Tintometer® Group



## Indice FAC (PFXi195)

Code: 139840



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Indice FAC (PFXi195)

A set of five glass conformance filters for checking PFXi instruments in FAC scale.

### Fiche technique

Glass Standard Values	Valeurs approximatives :
	- 7
	- 13
	- 15
	- 29
	- 39

Filters 5

### Contenu de livraison

- FAC 7 filter
- FAC 13 filter
- FAC 15 filter
- FAC 29 filter
- FAC 39 filter
- Certificate of conformity