

# Lovibond® Colour Measurement

## Tintometer® Group



### Lovibond® rouge, jaune, neutre (PFXi 995/950/880)

Code: 139630



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

#### Lovibond® rouge, jaune, neutre (PFXi 995/950/880)

A set of five glass conformance filters for checking PFXi instruments (PFXi 995/950/880) in Lovibond RYBN scale.

#### Fiche technique

**Glass Standard Values** Valeurs approximatives :

0,5R 1,3Y 0,1N  
1,3R 3,3Y 0,5N  
2,5R 8,6Y 0,1N  
4,0R 22Y 0,1N  
7,1R 51Y 0,5N

**Filters**

5

#### Contenu de livraison

- Lovibond RYBN 0.5R 1.3Y 0.1N filter
- Lovibond RYBN 1.3R 3.3Y 0.5N filter
- Lovibond RYBN 2.5R 8.6Y 0.1N filter
- Lovibond RYBN 4.0R 22Y 0.1N filter
- Lovibond RYBN 7.1R 51Y 0.5N filter
- Certificate of conformity