

# Lovibond® Colour Measurement Tintometer® Group



## Set di filtri unità IP

N. ordine: 137660



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Set di filtri unità IP

A set of four glass conformance filters for checking the PFXi instruments in IP (method B) units

### Dati tecnici

Glass Standard Values	Approximate values
	- 0.75
	- 1.25
	- 2.0
	- 3.5

Filters 4

## Fornitura

- IP 0.75 filter
- IP 1.25 filter
- IP 2.0 filter
- IP 3.5 filter
- Certificate of conformity