

# Lovibond® Colour Measurement Tintometer® Group



## Set di filtri Klett (filtro blu KS-42)

Solo per l'uso con PFXi 880, 950, 995

N. ordine: 139709



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Set di filtri Klett (filtro blu KS-42)

A set of five glass conformance filters for checking Klett scale on PFXi instruments.

#### Dati tecnici

Glass Standard Values	Approximate Values
	- 21, 66, 162, 318, 616

---

Filters	5
---------	---

### Fornitura

- Klett 21 filter
- Klett 66 filter
- Klett 162 filter
- Klett 318 filter
- Klett 616 filter
- Certificate of conformity