

Lovibond® Colour Measurement Tintometer® Group



Set di filtri numero di colore ASTM (PFXi195)

N. ordine: 139810



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Set di filtri numero di colore ASTM (PFXi195)

A set of three conformance filters to check the PFXi instruments in ASTM D1500 scale.

Dati tecnici

Glass Standard Values	Approximate Values
	- 0.5
	- 3.5
	- 5.0

Filters	3
---------	---

Fornitura

- 0.5
- 3.5
- 5.0
- Certificate of Conformity