

Lovibond® Colour Measurement Tintometer® Group



Beta karoten

Ürün No: 169520



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Beta karoten

A glass conformance filter to check the Model Fx range of instruments accuracy in measuring beta Carotene.

Teknik Veri

Glass Standard Values – 45 ppm at 10 mm path length

Filters 1

Teslimat Kapsam Metni

- 45ppm filter
- Certificate of Conformity