

Lovibond® Colour Measurement Tintometer® Group



ASTM renk endeksi

Ürün No: 341000



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

ASTM renk endeksi

A set of three glass filters to check the AF650 Petroleum Oils Comparator in ASTM D1500 scale.

Teknik Veri

Glass Standard Values	Approximate Values
	- 0.5
	- 3.5
	- 5.0

Filters 3

Teslimat Kapsam Metni

- 0.5
- 3.5
- 5.0
- Certificate of conformity