

# Lovibond® Colour Measurement Tintometer® Group



## CIE L\*a\*b\* (with non RCMSi PFXi195)

Part Number: 130530



Conformance filters are assigned a nominal value and the company always endeavors to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### CIE L\*a\*b\* (with non RCMSi PFXi195)

A single glass conformance filter for checking the accuracy of PFXi instruments in CIE xyY color space.

#### Technical Data

<b>Glass Standard Values</b>	Approximate values – L* 57 a* 24 b* 89
<b>Filters</b>	1

### Delivery Scope

- CIE filter
- Certificate of conformity