

Lovibond® Colour Measurement Tintometer® Group



IP 单位滤光片组

货号: 137660



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

IP 单位滤光片组

A set of four glass conformance filters for checking the PFXi instruments in IP (method B) units

Technical Data

Glass Standard Values	0.75 1.25 2.0 3.5
Filters	4

标准配置文本

- IP 0.75 filter
- IP 1.25 filter
- IP 2.0 filter
- IP 3.5 filter
- Certificate of conformity