

Lovibond® Colour Measurement Tintometer® Group



EP 棕色/黄色滤光片组

货号: 139440



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

EP 棕色/黄色滤光片组

A set of two glass conformance filters for checking PFXi instruments in European Pharmacopoeia Brown/Yellow scale.

Technical Data

Glass Standard Values	BY3, BY5
Filters	2

标准配置文本

- EP Brown/Yellow 3 filter
- EP Brown/Yellow 5 filter
- Certificate of Conformity