

Lovibond® Colour Measurement Tintometer® Group



CIE L*a*b* (with non RCMSi PFXi995/950/880)

Artigo No: 130520



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

CIE L*a*b* (with non RCMSi PFXi995/950/880)

A single glass conformance filter for checking the accuracy of PFXi instruments in CIE xyY colour space.

Dados Técnicos

Glass Standard Values - Valores aproximados:
- L* 57 a* 24 b*88

Filters 1

Texto do Escopo de Entrega

- CIE filter
- Certificate of conformity