

Lovibond® Colour Measurement Tintometer® Group



AOCS

Artigo No: 136980



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

AOCS

A single glass filter for testing the PFXi 880, 950 or 995 in AOCS-Tintometer scale.

Dados Técnicos

Glass Standard Values Valor aproximado: - 0.5R, 1.0Y

Filters 1

Texto do Escopo de Entrega

- 0.5R 1.0Y
- Certificate of Conformity