Iron PP / M222



Iron PP	M222
0.02 - 3 mg/L Fe ⁹⁾	FE1
1,10-Phenanthroline	

Instrument specific information

The test can be performed on the following devices. In addition, the required cuvette and the absorption range of the photometer are indicated.

Instrument Type	Cuvette	λ	Measuring Range
MD 100, MD 600, MD 610, MD 640, MultiDirect	ø 24 mm	530 nm	0.02 - 3 mg/L Fe ⁹⁾
XD 7000, XD 7500	ø 24 mm	510 nm	0.02 - 3 mg/L Fe ⁹⁾
SpectroDirect	□ 50 mm	510 nm	0.01 - 1.5 mg/L Fe ⁹⁾

Material

Required material (partly optional):

Reagents	Packaging Unit	Part Number
VARIO Ferro F10	Powder / 100 pc.	530560
VARIO Ferro F10	Powder / 1000 pc.	530563

Application List

- Waste Water Treatment
- Cooling Water
- Boiler Water
- · Galvanization
- Drinking Water Treatment
- Raw Water Treatment



Preparation


- 1. Iron oxide requires mild, strong or Digesdahl digestion before the analysis (digestion process with acid).
- 2. Very strong alkaline or acidic water samples should be adjusted to between pH 3 and pH 5 before the analysis.
- 3. Water samples containing visible rust should be allowed to react for at least five minutes.
- 4. Water that has been treated with organic compounds such as corrosion inhibitors, must be oxidised where necessary to break down the iron complex. 1 ml of concentrated Sulphuric acid (≥ 95 %) and 1 ml concentrated Nitric acid (≥ 65 %) is therefore added to to 100 ml water sample and boiled down to approximately half the volume. After cooling down, the digestion procedure is continued.

Notes

- 1. This method is for the determination of all forms of dissolved iron and most forms of undissolved iron.
- 2. Accuracy is not affected by undissolved powder.



Digestion



This sample is used for the analysis of total solved and dissolved Iron.



Determination of Iron (II,III), dissolved with Vario Powder Packs


Select the method on the device.

For testing of Iron with tablet, carry out the described digestion.

For this method, a ZERO measurement does not have to be carried out every time on the following devices: XD 7000, XD 7500







Fill 24 mm vial with **10 mL** Clos **sample**.

Close vial(s).

Place **sample vial** in the sample chamber. Pay attention to the positioning.





Press the ZERO button.

Remove the vial from the sample chamber.

For devices that require no ZERO measurement, start here.



Add Vario FERRO F10 powder pack.




Close vial(s).



Invert several times to mix the contents.





sample chamber. Pay attention to the positioning. START)button.

Wait for 3 minute(s) reaction time.

3 min

Once the reaction period is finished, the measurement takes place automatically.

The result in mg/L Iron appears on the display.



Chemical Method

1,10-Phenanthroline

Appendix

Calibration function for 3rd-party photometers

Conc. = $a + b \cdot Abs + c \cdot Abs^2 + d \cdot Abs^3 + e \cdot Abs^4 + f \cdot Abs^5$

	ø 24 mm	□ 10 mm	
а	-6.44557 • 10 ⁻²	-6.44557 • 10 ⁻²	
b	2.39506 • 10 ⁺⁰	5.14938 • 10 ⁺⁰	
С			
d			
е			
f			

Interferences

Persistant Interferences

1. Iridium interferes with the test.

According to

DIN 38406-E1 Standard Method 3500-Fe-1997 US EPA 40 CFR 136

⁹⁾ Reagent recovers most insoluble iron oxides without digestion