



Iron LR L (A)

M225

0.03 - 2 mg/L Fe

FE

Ferrozine / Thioglycolate

Instrument specific information

The test can be performed on the following devices. In addition, the required cuvette and the absorption range of the photometer are indicated.

Instrument Type	Cuvette	λ	Measuring Range
MD 100, MD 110, MD 600, MD 610, MD 640, Test Kit, XD 7000, XD 7500	ø 24 mm	560 nm	0.03 - 2 mg/L Fe

Material

Required material (partly optional):

Reagents	Packaging Unit	Part Number
Acidity / Alkalinity P Indicator PA1	65 mL	56L013565
Hardness Calcium Buffer CH2	65 mL	56L014465
KP962-Ammonium Persulphate Powder	Powder / 40 g	56P096240
KS63-FE6-Thioglycolate/Molybdate HR RGT	30 mL	56L006330
Iron Reagent FE6	65 mL	56L006365
Iron Reagent FE5	65 mL	56L006165


Application List

- Cooling Water
- Boiler Water
- Galvanization
- Raw Water Treatment




Preparation

1. If there are strong complexing agents in the sample, the response time must be extended until no further colour development is seen. However, very strong iron complexes are not included in the measurement. In this event, the complexing agent must be destroyed by means of oxidation with acid/persulphate and the sample also neutralised to pH 6–9.
2. For the measurement of total iron, both suspended and dissolved, the sample must be boiled with acid/persulphate. It must be neutralised back to pH 6–9 and refilled to the original volume with deionised water.




Digestion


Total iron consists of suspended, soluble and complexed iron. The sample must be not filtered before measuring. To ensure homogenisation of the sample, deposited particles must be evenly distributed immediately prior to sampling by forcible shaking. A filtration of the sample is necessary for the determination of total soluble iron (including the complex iron compounds). The equipment required for the determination of total iron and reagents are not included in the standard delivery.




Fill a suitable digestion vessel with **50 mL** homogenised sample.




Add **5 mL** 1:1 Hydrochloric acid.




Add a measuring scoop **KP 962 (Ammonium Persulfat Powder)**.




Boil the sample for **20 minutes**. A sample volume of about 25 mL should be retained; If necessary, fill with deionised water.




Allow the sample to cool to room temperature.




Hold cuvettes vertically and add equal drops by pressing slowly.



Add **1 drop Acidity / Alkalinity P Indicator PA1.**



Add **Hardness Calcium Buffer CH2** drop by drop to the same sample until colouration turns from light pink to red. **(Note: make sure to swirl the vial after adding each drop!)**



Fill the sample with **deionised water to 50 mL .**



Determination of Iron, total LR (A) with liquid reagent


Select the method on the device.

For testing of **Iron, total LR**, carry out the described **digestion**.


For this method, a **ZERO** measurement does not have to be carried out every time on the following devices: XD 7000, XD 7500




Fill 24 mm vial with **10 mL deionised water** .




Close vial(s).




Place **sample vial** in the sample chamber. Pay attention to the positioning.



Press the **ZERO** button.




Remove the vial from the sample chamber.




Empty vial.


For devices that require **no ZERO measurement** , **start here**.





Fill 24 mm vial with **10 mL prepared sample** .




Hold cuvettes vertically and add equal drops by pressing slowly.




Add **10 drops Iron Reagent FE5**.




Close vial(s).




Invert several times to mix the contents.



Place **sample vial** in the sample chamber. Pay attention to the positioning.




Press the **TEST** (XD: **START**) button.



Wait for **5 minute(s) reaction time**.

Once the reaction period is finished, the measurement takes place automatically.

The result in mg/L total Iron or when using a filtrated sample, in mg/l totale soluble Iron appears on the display.




Determination of Iron LR (A) with liquid reagent


Select the method on the device.

For this method, a ZERO measurement does not have to be carried out every time on the following devices: XD 7000, XD 7500


For determination of total dissolved iron the sample must be filtered prior to the test (pore size 0,45 µm). Otherwise, iron particles and suspended iron are measured.




Fill 24 mm vial with **10 mL prepared sample**.




Close vial(s).



Place **sample vial** in the sample chamber. Pay attention to the positioning.




Press the **ZERO** button.




Remove the vial from the sample chamber.


For devices that require **no ZERO measurement**, start here.




Hold cuvettes vertically and add equal drops by pressing slowly.




Add **10 drops Iron Reagent FE5**.




Close vial(s).




Invert several times to mix the contents.



Place **sample vial** in the sample chamber. Pay attention to the positioning.




Press the **TEST** (XD: **START**) button.



Wait for **5 minute(s) reaction time**.

Once the reaction period is finished, the measurement takes place automatically. The result in mg/L Iron appears on the display.



Chemical Method

Ferrozine / Thioglycolate

Appendix

Calibration function for 3rd-party photometers

Conc. = a + b•Abs + c•Abs² + d•Abs³ + e•Abs⁴ + f•Abs⁵

	∅ 24 mm	□ 10 mm
a	-2.05635 • 10 ⁻²	-2.05635 • 10 ⁻²
b	9.74475 • 10 ⁻¹	2.09512 • 10 ⁺⁰
c		
d		
e		
f		

Interferences

Removeable Interferences

1. If using KS61 (Ferrozine/Thioglycolate), a high concentration of molybdate will result in an intense yellow colour. In this instance, a chemical blank value is required:
 - Use two clean **24 mm vials**.
 - Mark one as blank for zeroing.
 - Fill a clean vial (24 mm) with **10 ml of the sample** (blank).
 - Add **10 drops of KS63 (Thioglycolate)** to the vial.
 - Close the vial with the cap and swirl the contents to mix them.
 - Place the blank in the sample chamber.
 - Pay attention to the positioning.
 - Press the **ZERO** button.
 - Remove the vial from the sample chamber.
 - Fill a second clean vial (24 mm) with **10 ml of the sample** (this is the sample vial).
 - Add **10 drops of KS63 (Ferrozine/Thioglycolate)** and as before, follow the procedure as described.



Interference	from / [mg/L]
Co	8
Cu	2
Oxalat	500
CN ⁻	10
NO ₂ ⁻	

Bibliography

D. F. Boltz and J. A. Howell, eds., *Colorimetric Determination of Nonmetals*, 2nd ed., Vol. 8, p. 304 (1978). Carpenter, J.F. "A New Field Method for Determining the Levels of Iron Contamination in Oilfield Completion Brine", SPE International Symposium (2004)