

# Lovibond® Colour Measurement

## Tintometer® Group



## Lovibond Red, Yellow, Neutral Filter Set (PFXi 995/950/880)

Part Number: 139610



Conformance filters are assigned a nominal value and the company always endeavors to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Lovibond Red, Yellow, Neutral Filter Set (PFXi 995/950/880)

A set of five glass conformance filters for checking PFXi instruments in Lovibond RYBN scale.

#### Technical Data

Glass Standard Values	Approximate Values
	- 0.1R 0.5Y 0.0N
	- 1.0R 2.6Y 0.1N
	- 2.0R 5.7Y 0.1N
	- 3.2R 15.0Y 0.1N
	- 5.0R 34.0Y 0.1N

Filters

5

#### Delivery Scope

- Lovibond 0.1R 0.5Y 0.0N filter
- Lovibond 1.0R 2.6Y 0.1N filter
- Lovibond 2.0R 5.7Y 0.1N filter
- Lovibond 3.2R 15.0Y 0.1N filter
- Lovibond 5.0R 34.0Y 0.1N filter
- Certificate of conformity