


**Conductivity****SD / Con****0 - 20 mS/cm****Application List**

- Cooling Water


**Notes**

1. The description of the calibration and the device settings are described in the detailed operating instructions. Detailed operating instructions are enclosed with the device.






Rinse the electrode with distilled or deionised water and carefully wipe with a paper towel.



Hold the tester without protective cap in the sample water to be measured so that the electrode and temperature probe is immersed in the sample water not deeper than the seal ring.



The measurement result is displayed.