

Lovibond® Colour Measurement Tintometer® Group



Lot de filtres AOCS-Tintometer (Fx)

Code: 166900



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Lot de filtres AOCS-Tintometer (Fx)

A set of three filters for testing the Model Fx in AOCS-Tintometer scale.

Fiche technique

Glass Standard Values	Valeurs approximatives :
	0,5R 1,0Y,
	3,7R 13,0Y,
	6,6R 50Y

Filters	3
----------------	---

Contenu de livraison

- 0.5R 1.0Y
- 3.7R 13.0Y
- 6.6R 50Y
- Certificate of Conformity