

# Lovibond® Colour Measurement Tintometer® Group



## CIE xyY

Code: 136970



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### CIE xyY

A single glass conformance filter for checking the accuracy of colour instruments in CIE xyY colour space.

### Fiche technique

**Glass Standard Values** - Valeurs approximatives :  
- x0.47 y0.45 y44.5

**Filters** 1

## Contenu de livraison

- CIE filter
- Certificate of conformity