

Lovibond® Colour Measurement Tintometer® Group



Lot de filtres Ph.Eur vert/jaune

Code: 139450



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Lot de filtres Ph.Eur vert/jaune

A set of two glass conformance filters for checking PFXi instruments in European Pharmacopoeia Green/Yellow scale.

Fiche technique

Glass Standard Values GY2, GY5

Filters 2

Contenu de livraison

- EP Green/Yellow 2 filter
- EP Green/Yellow 5 filter
- Certificate of Conformity