

Lovibond® Colour Measurement Tintometer® Group



Lot de filtres Lovibond® rouge, jaune, neutre (PFXi 995/950/880)

Code: 139610



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Lot de filtres Lovibond® rouge, jaune, neutre (PFXi 995/950/880)

A set of five glass conformance filters for checking PFXi (PFXi 995/950/880) instruments in Lovibond RYBN scale.

Fiche technique

| Glass Standard Values | Valeurs approximatives : |
|-----------------------|--------------------------|
| | - 0.1R 0.5Y 0.0N |
| | - 1.0R 2.6Y 0.1N |
| | - 2.0R 5.7Y 0.1N |
| | - 3.2R 15.0Y 0.1N |
| | - 5.0R 34.0Y 0.1N |

Filters

5

Contenu de livraison

- Lovibond 0.1R 0.5Y 0.0N filter
- Lovibond 1.0R 2.6Y 0.1N filter
- Lovibond 2.0R 5.7Y 0.1N filter
- Lovibond 3.2R 15.0Y 0.1N filter
- Lovibond 5.0R 34.0Y 0.1N filter
- Certificate of conformity