

Lovibond® Colour Measurement Tintometer® Group



Lot de filtres selon le livre médical US- américain (US Pharmacopoeia)

Code: 139460



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Lot de filtres selon le livre médical US-américain (US Pharmacopoeia)

A set of three glass conformance filters for checking PFXi instruments in US Pharmacopoeia scale.

Fiche technique

Glass Standard Values G, H, P

Filters 3

Contenu de livraison

- USP G filter
- USP H filter
- USP P filter
- Certificate of conformity