

Lovibond® Colour Measurement Tintometer® Group



Número de cor do AOCS Tintometer (AF 710-3)

Artigo No: 109720



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Número de cor do AOCS Tintometer (AF 710-3)

A set of three filters for testing the accuracy of the AF710-3 in AOCS-Tintometer scale.

Dados Técnicos

Glass Standard Values	0.5R 1.0Y, 3.7R 13.0Y, 6.6R 38Y.
------------------------------	--

Filters	3
----------------	---

Texto do Escopo de Entrega

- 0.5R, 1.0Y
- 3.7R, 13.0Y
- 6.6R, 38Y
- Certificate of Conformity