



Alkalinity-m T

M30

5 - 200 mg/L CaCO<sub>3</sub>

tA

Acid / Indicator

## Instrument specific information

The test can be performed on the following devices. In addition, the required cuvette and the absorption range of the photometer are indicated.

Instrument Type	Cuvette	$\lambda$	Measuring Range
MD 100, MD 110, MD 200, MD 600, MD 610, MD 640, MultiDirect, PM 600, PM 620, PM 630	$\varnothing$ 24 mm	610 nm	5 - 200 mg/L CaCO <sub>3</sub>
SpectroDirect, XD 7000, XD 7500	$\varnothing$ 24 mm	615 nm	5 - 200 mg/L CaCO <sub>3</sub>

## Material

Required material (partly optional):

Reagents	Packaging Unit	Part Number
Alka-M-Photometer	Tablet / 100	513210BT
Alka-M-Photometer	Tablet / 250	513211BT

## Application List

- Drinking Water Treatment
- Waste Water Treatment
- Raw Water Treatment
- Pool Water Control

## Notes

1. The terms Alkalinity-m, m-Value, total alkalinity and Acid demand to K<sub>S4,3</sub> are identical.
2. For accurate results, exactly 10 ml of water sample must be used for the test.






## Determination of Alkalinity, total = Alkalinity-m = m-Value with Tablet

Select the method on the device.


For this method, a ZERO measurement does not have to be carried out every time on the following devices: XD 7000, XD 7500



Fill 24 mm vial with **10 mL** sample.

Close vial(s).


Place **sample vial** in the sample chamber. Pay attention to the positioning.



Press the **ZERO** button.

Remove the vial from the sample chamber.



For devices that require **no ZERO measurement**, start here.




Add **ALKA-M-PHOTOMETER** tablet.

Crush tablet(s) by rotating slightly.


Close vial(s).



Dissolve tablet(s) by inverting.



Place **sample vial** in the sample chamber. Pay attention to the positioning.



Press the **TEST (XD: START)**button.

The result in Alkalinity-m appears on the display.



## Analyses

The following table identifies the output values can be converted into other citation forms.

Unit	Cite form	Scale Factor
mg/l	CaCO <sub>3</sub>	1
	°dH	0.056
	°eH	0.07
	°fH	0.1
	°aH	0.058
	K <sub>S4.3</sub>	0.02

## Chemical Method

Acid / Indicator

## Appendix

### Calibration function for 3rd-party photometers

$$\text{Conc.} = a + b \cdot \text{Abs} + c \cdot \text{Abs}^2 + d \cdot \text{Abs}^3 + e \cdot \text{Abs}^4 + f \cdot \text{Abs}^5$$

	ø 24 mm	□ 10 mm
a	-2.46587 • 10 <sup>+1</sup>	-2.46587 • 10 <sup>+1</sup>
b	2.67915 • 10 <sup>+2</sup>	5.76017 • 10 <sup>+2</sup>
c	-1.48158 • 10 <sup>+2</sup>	-6.84858 • 10 <sup>+2</sup>
d	5.11097 • 10 <sup>+1</sup>	5.07947 • 10 <sup>+2</sup>
e		
f		

### Derived from

EN ISO 9963-1