

Phosphate LR L

M334

0.1 - 10 mg/L PO<sub>4</sub>

Phosphomolybic Acid / Ascorbic Acid

## Instrument specific information

The test can be performed on the following devices. In addition, the required cuvette and the absorption range of the photometer are indicated.

Instrument Type	Cuvette	$\lambda$	Measuring Range
MD 600, MD 610, MD 640, XD 7000, XD 7500	ø 24 mm	660 nm	0.1 - 10 mg/L PO <sub>4</sub>

## Material

Required material (partly optional):

Reagents	Packaging Unit	Part Number
KS278-Sulphuric Acid 50 % V/V	65 mL	56L027865
Acidity / Alkalinity P Indicator PA1	65 mL	56L013565
Hardness Calcium Buffer CH2	65 mL	56L014465
KP962-Ammonium Persulphate Powder	Powder / 40 g	56P096240
Phosphate LR Reagent Pack	1 pc.	56R023765

## Application List

- Waste Water Treatment
- Boiler Water
- Drinking Water Treatment
- Raw Water Treatment
- Pool Water Control

## Preparation

1. Strongly buffered samples or samples with extreme pH values should be adjusted to between pH 6 and pH 7 before the analysis (use 1 mol/l Sulphuric acid or 1 mol/l Sodium hydroxide).
2. Prior digestion is required for the analysis of Polyphosphate and total phosphate.




## Notes


1. The measuring spoon supplied with the reagents must be used for the correct dosage.
2. The long scoop is for KP962. The short scoop is for KP119.




## Digestion total Phosphate LR with liquid reagents




Fill a suitable digestion vessel with **50 mL** homogenised sample.




Add a measuring scoop **KP962 (Ammonium Persulfate Powder)**.




Add **15 drops KS278 (50% sulfuric acid)**.




Boil the sample for **20 minutes**. A sample volume of about 25 mL should be retained; If necessary, fill with deionised water.




Invert the vial and allow to cool to room temperature.



Add **2 drops Acidity / Alkalinity P Indicator PA1**.




Add **Hardness Calcium Buffer CH2** drop by drop to the same sample until colouration turns from light pink to red. **(Note: make sure to swirl the vial after adding each drop!)**




Fill the sample with **deionised water to 50 mL**.







## Digestion Polyphosphate LR with liquid reagents




Fill a suitable digestion vessel with **50 mL homogenised sample**.




Add **15 drops KS278 (50% sulfuric acid)**.




Boil the sample for **20 minutes**. A sample volume of about 25 mL should be retained; If necessary, fill with deionised water.




Invert the vial and allow to cool to room temperature.



Add **2 drops Acidity / Alkalinity P Indicator PA1**.




Add **Hardness Calcium Buffer CH2** drop by drop to the same sample until colouration turns from light pink to red. **(Note: make sure to swirl the vial after adding each drop!)**



Fill the sample with **deionised water to 50 mL**.







## Determination of Phosphate LR with liquid reagent

Select the method on the device.


For this method, a ZERO measurement does not have to be carried out every time on the following devices: XD 7000, XD 7500




Filter approx. 14 mL sample with a pre-rinsed filter (pore size 0.45 µm).




Fill 24 mm vial with **10 mL prepared sample**.




Close vial(s).



Place **sample vial** in the sample chamber. Pay attention to the positioning.




Press the **ZERO** button.




Remove the vial from the sample chamber.


For devices that require **no ZERO measurement**, start here.




Hold cuvettes vertically and add equal drops by pressing slowly.




Add **50 drops KS80 (CRP)**




Close vial(s).




Invert several times to mix the contents.




Add a measuring scoop **KP119 (Ascorbic Acid)**.




Close vial(s).




Swirl around to dissolve the powder.



Place **sample vial** in the sample chamber. Pay attention to the positioning.



Press the **TEST** (XD: **START**) button.



Wait for **10 minute(s) reaction time**.

Once the reaction period is finished, the measurement takes place automatically. The result in mg/L Phosphate appears on the display.





## Determination of Polyphosphate LR with liquid reagents

Select the method on the device.

For this method, a ZERO measurement does not have to be carried out every time on the following devices: XD 7000, XD 7500

For testing of **Polyphosphate LR with liquid reagents**, carry out the described **digestion**.

This test determines the content of inorganic total phosphate. The Polyphosphate content arises from the difference between inorganic and ortho phosphate.

The test for Polyphosphate LR with liquid reagents runs just as the test under Method 334, Phosphate LR with liquid reagents.

The result in mg/L anorganic Total Phosphate (ortho-Phosphate and Polyphosphate) appears on the display.





## Determination of total Phosphate LR with liquid reagent

Select the method on the device.

For testing of **total Phosphate LR with liquid reagents**, carry out the described **digestion**.

This test determines all compounds of phosphorus present in the sample, including ortho-phosphate, polyphosphate, and organic phosphorus compounds.

The test for total Phosphate LR with liquid reagents runs just as the test under Method 334, Phosphate LR with liquid reagents.

The result in mg/L total Phosphate appears on the display.

## Analyses

The following table identifies the output values can be converted into other citation forms.

Unit	Cite form	Scale Factor
mg/l	P	1
mg/l	PO <sub>4</sub> <sup>3-</sup>	3.066177
mg/l	P <sub>2</sub> O <sub>5</sub>	2.29137

## Chemical Method

Phosphomolybic Acid / Ascorbic Acid

## Appendix

### Calibration function for 3rd-party photometers

Conc. = a + b•Abs + c•Abs<sup>2</sup> + d•Abs<sup>3</sup> + e•Abs<sup>4</sup> + f•Abs<sup>5</sup>

	∅ 24 mm	□ 10 mm
a	-4.14247 • 10 <sup>-2</sup>	-4.14247 • 10 <sup>-2</sup>
b	1.33552 • 10 <sup>+0</sup>	2.87137 • 10 <sup>+0</sup>
c	-2.89775 • 10 <sup>-1</sup>	-1.33948 • 10 <sup>+0</sup>
d	2.04577 • 10 <sup>-1</sup>	2.03316 • 10 <sup>+0</sup>
e		
f		

## Interferences

### Persistent Interferences

- Large amounts of unresolved substances can cause non-reproducible measurement results.



<b>Interference</b>	<b>from / [mg/L]</b>
Al	200
AsO <sub>4</sub> <sup>3-</sup>	in all quantities
Cr	100
Cu	10
Fe	100
Ni	300
SiO <sub>2</sub>	50
Si(OH) <sub>4</sub>	10
S <sup>2-</sup>	in all quantities
Zn	80

**According to**

DIN ISO 15923-1 D49

Standard Method 4500-P E

US EPA 365.2