

Lovibond® Colour Measurement Tintometer® Group



Saybolt - EC3000

Part Number: 135160



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Saybolt - EC3000

A set glass conformance filters for checking the EC3000 instrument in Saybolt scale.

Technical Data

Glass Standard Values	Approximate Value:
	- +20
	- +0
	- Zero baseline

Filters	3
---------	---

Delivery Scope

- Saybolt, +20, +0, Zero baseline filters
- Certificate of conformity