

Lovibond® Colour Measurement Tintometer® Group



Filterset AOCs-Tintometer (PFXi 995/950/880)

Bestelnr.: 136900



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Filterset AOCs-Tintometer (PFXi 995/950/880)

A set of three filters for testing the PFXi 880, 950 or 995 in AOCs-Tintometer scale.

Technische gegevens

Glass Standard Values	Approximate Values
	- 0.5R 1.0Y,
	- 3.7R 13.0Y,
	- 6.6R 50Y

Filters 3

Leveromvang

- 0.5R 1.0Y
- 3.7R 13.0Y
- 6.6R 50Y
- Certificate of Conformity