

Lovibond® Colour Measurement Tintometer® Group



Klorofil A (Fx)

Ürün No: 169530



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Klorofil A (Fx)

A glass conformance filter to check the Model Fx range of instruments accuracy in measuring Chlorophyll A.

Teknik Veri

Glass Standard Values – 0.19 ppm at 10 mm path length

Filters 1

Teslimat Kapsam Metni

- 0.19 ppm filter
- Certificate of Conformity