

# Lovibond® Colour Measurement Tintometer® Group



## AOCS-Tintometer renk endeksi (AF 710-3)

Ürün No: 109720



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### AOCS-Tintometer renk endeksi (AF 710-3)

A set of three filters for testing the accuracy of the AF710-3 in AOCS-Tintometer scale.

#### Teknik Veri

<b>Glass Standard Values</b>	- 0.5R 1.0Y
	- 3.7R 13.0Y
	- 6.6R 38Y

Filters 3

### Teslimat Kapsam Metni

- 0.5R, 1.0Y
- 3.7R, 13.0Y
- 6.6R, 38Y
- Certificate of Conformity