

# Lovibond® Colour Measurement Tintometer® Group



## Lot de filtres à indice Gardner (PFXi195)

Code: 139880



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Lot de filtres à indice Gardner (PFXi195)

A set of four glass conformance filters for checking PFXi instruments in Gardner scale.

#### Fiche technique

<b>Glass Standard Values</b>	Valeurs approximatives :
	2
	8
	12
	17
<b>Filters</b>	4

### Contenu de livraison

- Gardner 2 filter
- Gardner 8 filter
- Gardner 12 filter
- Gardner 17 filter
- Certificate of conformity