

Lovibond® Colour Measurement Tintometer® Group



Lot de filtres, indice Hazen (uniquement PFX/ PFXi 880/950/995)

Code: 135150



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Lot de filtres, indice Hazen (uniquement PFX/PFXi 880/950/995)

Glass conformance filters for checking Pt-Co scale on EC 2000 PT-Co.

Fiche technique

Glass Standard Values	Valeurs approximatives :
	- 65
	- Zero Baseline

Filters	2
----------------	---

Contenu de livraison

- 65, Zero Baseline filters
- Certificate of Conformity