

Lovibond® Colour Measurement Tintometer® Group



Gardner - EC2000

Code: 135139



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Gardner - EC2000

A single glass conformance filter for checking EC2000 instrument in-Gardner scale.

Fiche technique

Glass Standard Values - Valeurs approximatives :
es - 11.3

Filters 1

Contenu de livraison

- Gardner 11.3 filter
- Certificate of Conformity