

# Lovibond® Colour Measurement Tintometer® Group



## Lot de filtres, indice de Saybolt (uniquement PFX/PFXi 195)

Code: 139390



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Lot de filtres, indice de Saybolt (uniquement PFX/PFXi 195)

A set of five glass conformance filters for checking the PFXi 195 instruments in Saybolt scale. NOT FOR USE with the PFXi 880, 850 or 995.

### Fiche technique

**Glass Standard Values** -8, 0, +10, +18, +25

**Filters** 5

### Contenu de livraison

- Saybolt -8 filter
- Saybolt 0 filter
- Saybolt +10 filter
- Saybolt +18 filter
- Saybolt +25 filter
- Certificate of conformity