

Lovibond® Colour Measurement Tintometer® Group



Parafuso - EC3000

Artigo No: 135160



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Parafuso - EC3000

A set glass conformance filters for checking the EC3000 instrument in Saybolt scale.

Dados Técnicos

Glass Standard Values	Valor aproximado: - - +20 - +0 - Zero baseline
------------------------------	---

Filters	3
----------------	---

Texto do Escopo de Entrega

- Saybolt, +20, +0, Zero baseline filters
- Certificate of conformity