

Lovibond® Colour Measurement Tintometer® Group



Número de cor FAC

Artigo No: 343000



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

Número de cor FAC

A set of four glass conformance filters for checking the 3000 Comparator in FAC scale

Dados Técnicos

Glass Standard Values	Valor aproximado: - 7, 13, 29, 39
Filters	4

Texto do Escopo de Entrega

- FAC 7 filter
- FAC 13 filter
- FAC 29 filter
- FAC 39 filter
- Certificate of conformity