

# Lovibond® Colour Measurement Tintometer® Group



## Conjunto de filtros US pharmacopoeia (farmacopeia americana)

Artigo No: 139460



Conformance filters are assigned a nominal value and the company always endeavours to match this value as closely as possible. On occasion, because of the unique manner of the glass making process, it may not be possible to match the requested values exactly. The values achieved and the performance tolerances expected will, of course, always be reported.

### Conjunto de filtros US pharmacopoeia (farmacopeia americana)

A set of three glass conformance filters for checking PFXi instruments in US Pharmacopoeia scale.

#### Dados Técnicos

<b>Glass Standard Values</b>	G, H, P
<b>Filters</b>	3

### Texto do Escopo de Entrega

- USP G filter
- USP H filter
- USP P filter
- Certificate of conformity